ORCHIDOMETER: USER MANUAL

INTRODUCTION:

An orchidometer is used to measure the volume of the testicles. It consists of a string of twelve numbered wooden or plastic beads of increasing size from about 1 to 25 milliliters. The orchidometer can be used to determine size of testes. Discrepancy of testicular size with other parameters of maturation can be an important clue to various diseases.

SYMBOLS:

Symbol	Definition of Symbol
	ATTENTION: See Instructions for Use
	CE mark has been identified according to Article 17 of the MDD(93/42/EEC)
	Symbol for "AUTHORISED REPRESENTATIVE IN THE EUROPEAN COMMUNITY"
	Symbol for "BATCH CODE"
	This symbol shall be accompanied by the name and the address of the manufacturer
	Symbol for "TEMPERATURE LIMITATION"
	This symbol shall be accompanied by a date to indicate that the device should not be used after the end of the year, month or day shown.
	Symbol for "CATALOGUE NUMBER"

PLEASE READ THE FOLLOWING INFORMATION BEFORE USING THE ORCHIDOMETER:

Warnings

1. Keep it out of children's reach.
2. Keep it away from flame.

Caution

1. It is necessary to read this user's manual carefully before using this product.
2. Follow local governing ordinances and recycling instructions regarding disposal or recycling of the device and components.
3. Please handle orchidometer with care. If there is any damage resulting from abnormal use, please contact service personnel.

CLEANING AND MAINTENANCE:

Use a cleaning cloth or a tissue paper to clean and maintain the surface of the Orchidometer.

INSTRUCTIONS FOR USING THE ORCHIDOMETER

In order to measure the volume of the testicles of a patient, beads of Orchidometer are compared to the testicle area. The volume number on the bead that best matches the size of a testicle is then read.

Number on the bead that best matches the size of a testicle	Indicated Sexual Development Stage
1~3	Childhood
4, 6	Early Puberty
8, 10	Mid Puberty
14, 16	Late Puberty
18, 25	Adulthood

LIST OF ACCESSORIES AND DETACHABLE PARTS

Item	Number	Description	Quantity
Main device	A474-001	Main device	1 set (12 beads)
Metal tag	A474-002	For branding purpose	1 pc
Paper box	A474-003	White or printed paper box as the packaging of the device	1 pc

MISCELLANEOUS

Technical support

If you are not able to solve a problem, please contact your retail seller. You can also contact our technical support personnel by phone or mail, according to the following contact information:
Service call: +886 (02)25319955
Address: 9F, No.129 Zhong Shan N Road Sec.2, Taipei, Taiwan

Warranty

APPLE BioMedical Inc. provides a one-year warranty, starting from the purchasing date, to the original purchaser for repairing any product failure due to fabrication error or defective parts. Within the limited range of the warranty, defective products shall be fixed or replaced by APPLE BioMedical Inc. This warranty is not transferable. Damage due to negligence, accident, abuse, misapplication, modification, mailing to the seller, or repairs not made by APPLE BioMedical Inc., is not covered by the warranty. Within the coverage of the warranty, the delivery fee for mailing to the retail seller or the direct store is not covered within the scope of the warranty.

Reproduction or reprinting of part or all of the contents of this manual by any means without written authorization from APPLE BioMedical Inc.. is prohibited.

EC REP

Medical Device Safety Service GmbH
Schiffgraben 41, 30175 Hannover
Germany

APPLE BioMedical Inc.
9F, No.129 Zhong Shan N Road Sec.2, Taipei, Taiwan

APPLE BioMedical Inc.
9F, No.129 Zhong Shan N Road Sec.2, Taipei, Taiwan.
Telephone: +886 (02)25319955 Fax: +886 (02)25819922
Website: www.appleBMI.com

CE

A474-UM-001 01/10